

REPORT ON BIO-MEDICAL WASTE MANAGEMENT FOR THE MONTH OF MAY, 2021

Sl. No.	Description	Remarks
1	Contact Address and details of the Healthcare Facility	LGB Regional Institute of Mental Health, Tezpur, Sonitpur, Assam, Pin-784001, Tele No. 03712-232652, fax No.- 03712-233623
2	No. of Beds	336 No.
3	Details of :	Applied for, Application no: 444804 dated, 12/03/2021 Previous authorization No: WB/OTWA/BMW-448/18-19/01/170
	a) Authorisation under BMW Rules, 2016	Applied for, Application no: 444804, dated, 12/03/2021 Previous authorization No: WB/TEZ/T-1745/18-19/01/1489
4	b) Consent under Water (Prevention and Control of Pollution) Act, 1974 and Air(Prevention and Control of Pollution) Act, 1981	5.59 Kg
5	Quantity of bio-medical waste generation (in Kg/day)	Through captive treatment facility
6	Mode of disposal of bio-medical waste (through CBWTF or through captive treatment facility)	N/A
7	Name and address of the CBWTF through which waste is disposed off (as applicable)	
	In case, HCF is having captive treatment facility,	5.59 Kg
	a) Bio-medical waste treated (in Kg/day)	i) Incinerator ii) Autoclave iii) Shredder iv) Needle Destroyer v) ETP (Effluent Treatment Plant)
	b) Details of treatment equipment	i) Incinerator- 1 No. Capacity: 10-12 Kg/hr ii) Autoclave - 1 No. Capacity : 78 litres iii) Shredder - 1 No. Capacity : 50 -100 kg iv) Needle Destroyer - 12 Nos. — v) ETP - 1 No Capacity: 4000 litre/day
	c) Total nos. And capacity of each treatment equipment (in Kg/day)	

	<p>d) Operating parameters of the treatment equipment as per BMWWM Rules, 2016</p>	<p>1) Incinerator</p> <table border="1" data-bbox="1230 1346 1449 2067"> <thead> <tr> <th>Sl. No.</th> <th>Parameter</th> <th>Result</th> </tr> </thead> <tbody> <tr> <td>i.</td> <td>SO₂ (ppm)</td> <td>17.5</td> </tr> <tr> <td>ii.</td> <td>NO₂ (ppm)</td> <td>37.5</td> </tr> <tr> <td>iii.</td> <td>HCl (mg/Nm³)</td> <td>29</td> </tr> <tr> <td>iv.</td> <td>CO (ppm)</td> <td>286</td> </tr> <tr> <td>v.</td> <td>CO₂ %</td> <td>8.2</td> </tr> </tbody> </table> <p>Temperature Primary Chamber 800° ± 50° C, Secondary Chamber 1000° ± 50° C Height- 30 meters above from the ground Capacity- 10-12 Kg/hr</p> <p>2) Autoclave: (Gravity flow)</p> <p>i) Monitors steam processes with Geobacillus stearothermophilus spores. ii) Steam Indicator tape</p>	Sl. No.	Parameter	Result	i.	SO ₂ (ppm)	17.5	ii.	NO ₂ (ppm)	37.5	iii.	HCl (mg/Nm ³)	29	iv.	CO (ppm)	286	v.	CO ₂ %	8.2
Sl. No.	Parameter	Result																		
i.	SO ₂ (ppm)	17.5																		
ii.	NO ₂ (ppm)	37.5																		
iii.	HCl (mg/Nm ³)	29																		
iv.	CO (ppm)	286																		
v.	CO ₂ %	8.2																		
8	<p>Monthly records of biomedical waste generation(category wise)</p>	<p>i) Yellow category : 64.831 kg/month ii) Red category : 93.494 kg/month iii) White category : 2.397 kg/month iv) Blue category : 6.704 kg/month</p>																		
9	<p>No. of trainings conducted on Bio-medical Waste management in the current year</p>	<p>08</p>																		